

300 Stewart Road Wilkes-Barre, PA 18706 USA Phone: (570) 824-4666 Web: www.ipdpower.com Email: sales@ipdpower.com Fax: (570) 824-4843

**OUTPUT 4** 

 $-12V/2A_{(1)}$ 

-12V/2A<sub>(1)</sub>

 $-15V/2A_{(1)}$ 

-12V/2A<sub>(1)</sub>

 $-15V/2A_{(1)}$ 

-12V/2A<sub>(1)</sub> -15V/2A<sub>(1)</sub>

-12V/2A -15V/2A

## **DC4-70 SERIES OPERATING INSTRUCTIONS**

**INPUT RATINGS:** 36-72VDC, 4A

OUTPUT RATINGS: 70 Watts Maximum Total Continuous Output Power with 300 LFM Forced Air. Open Frame or with Chassis and Cover.

50 Watts Maximum Total Continuous Output Power Convection Cooled. Open Frame or with Chassis and Cover.

 $\frac{OUTPUT \ 3}{+12V/2A_{(1)}}$ 

 $+12V/2A_{(1)}$ 

 $+15V/2A_{(1)}$ 

 $+12V/2A_{(1)}$ 

 $+15V/2A_{(1)}$ 

 $+12V/2A_{(1)}$ 

 $+15V/2A_{(1)}$ 

40 Watts Maximum Total Continuous Output Power Convection Cooled with Chassis and Cover.

MODEL LISTING:

<u>MODEL</u>	OUTPUT 1	OUTPUT
DC4-70-4001	+3.3V/6A	+5V/5A
DC4-70-4002	+5V/6A	+3.3V/5A
DC4-70-4003	+5V/6A	+3.3V/5A
DC4-70-4004	+5V/6A	-5V/5A
DC4-70-4005	+5V/6A	-5V/5A
DC4-70-4006	+5V/6A	+24V/2A
DC4-70-4007	+5V/6A	+24V/2A
DC4-70-3001	+5V/6A	+12V/2A
DC4-70-3002	+5V/6A	+15V/2A
DC4-70-2001	+3.3V/6A	+5V/5A
DC4-70-2002	+5V/6A	+12V/4A
DC4-70-2003	+5V/6A	+24V/2A
DC4-70-2004	+12V/3A	-12V/3A
DC4-70-2005	+15V/3A	-15V/2A
DC4-70-2007	5V/6A	29V/2A
DC4-70-1001	2.5V/14A	
DC4-70-1002	3.3V/14A	
DC4-70-1003	5V/14A	
DC4-70-1004	12V/5.8A	
DC4-70-1005	15V/4.7A	
DC4-70-1006	24V/2.9A	
DC4-70-1007	28V/2.5A	
DC4-70-1008	48V/1.5A	

NOTES: 1. Rated 1.5 Amps with convection cooling.

A suffix may be added to the model number to indicate the following optional configurations: (CH-chassis, CO-cover, IO-isolated outputs, TS-terminal strip, BD- Blocking Diode).

# CE

## **DECLARATION OF CONFORMITY**



Manufacturer: Integrated Power Designs, Inc.

Manufacturer's Address: 300 Stewart Road, Wilkes-Barre, PA 18706 USA

Product Type: Switchmode DC-DC Power Supply

Declares all models listed above including all options are in compliance with the following European Community Directives:

## RoHS Directive 2015/863/EU of 31 March 2015

As well as with the requirements set out in the Laws of the United Kingdom relating to and applying the following standards:

### Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 SI No. 3032 + 2019 SI No.492

In addition, all models are Certified to be in compliance with applicable requirements of UL 62368-1:2014, ANSI/AAMI ES60601:2005/(R) 2012, IEC 62368-1:2014 and IEC 60601-1:2005/A1:2012 including all EU national deviations, CAN/CSA-C22.2 No. 62368-1-14, CAN/CSA-C22.2 No. 60601-1:2014, EN 62368-1:2014 and EN 60601-1:2006/A1:2013.

BY: Steven Thompson- President

PLACE: Integrated Power Designs

300 Stewart Road, Wilkes-Barre, PA 18706 USA

**DATE:** February 16, 2021

**EUROPEAN CONTACT:** 

Compumess Elektronik GmbH Lise-Meitner-Strasse 1 85716 Unterschleiβheim Telephone (089) 32 15 01-0 **UK CONTACT:** 

ALL PSU LTD Laser Quay, Culpeper Close Rochester Kent ME2 4HU Tel: 01634 725527

CLASSIFICATION:

- 1. Protection against electric shock Class I.
- 2. Protection against harmful ingress of water IPXO (Non-protected), ordinary.
- 3. Methods of sterilization None.
- 4. Suitability for use in an oxygen rich environment End user responsibility, not evaluated.
- 5. Mode of operation Continuous.



#### 300 Stewart Road Wilkes-Barre, PA 18706 USA Phone: (570) 824-4666

Web: www.ipdpower.com Email: sales@ipdpower.com Fax: (570) 824-4843

**ENVIRONMENTAL:** For Indoor Use Only.

No protective packaging required during transport.

Operating Temperature- 0-50°C Storage Temperature- -40-85°C

Operating / Storage Humidity- 0-95%RH non-condensing

Operating Altitude- 3000m max.

WARNING! SHOCK HAZARD: Dangerous voltages are present on some components, printed circuit board traces and heatsinks.

WARNING!

An open internal fuse indicates a catastrophic failure of circuit component(s). Repair must be by authorized IPD personnel only.

**RISK OF FIRE:** Refer to fuse rating on power supply circuit board for rating.

**WARNING!** Modification of this equipment is prohibited without the authorization of the manufacturer.

**SEPARATION:** Primary to secondary creepage distance is 6mm minimum, clearance 3.2mm minimum (2MOOP). Primary to ground creepage

distance is 3mm minimum, clearance 1.6mm minimum (1MOOP). Secondary to ground creepage is 2.3mm minimum, clearance is 1.4mm minimum (Operational Insulation). The required separation must be maintained in the end use equipment to preserve the

established means of protection.

OUTPUTS: The outputs are not acceptable for patient connection without additional isolation. All outputs are ES1 under normal and single

fault conditions unless otherwise indicated.

**TEMPERATURES:** The maximum operating temperatures of safety components as defined in the applicable safety standards must not be exceeded

after installation in the end use equipment. Output power, ambient air temperature and convection or forced air cooling availability

should be considered in the end use equipment.

HIPOT: In consideration of IEC 60601:2005 Clause 8.8.3, care must be taken to insure the voltage applied to a reinforced insulation does


not overstress different types and levels of insulation. Breakdown of basic insulation and catastrophic failure of the power supply may result if a test voltage of greater than 1500 VAC is applied between primary and secondary circuits. Each isolating component is factory tested at 3000 VAC minimum prior to installation. Primary and secondary to ground capacitors may need to be

disconnected prior to performing a dielectric strength test on the power supply or the end product.

**INSTALLATION:** The power supplies included in model listing on reverse side are considered components intended for professional installation into

end use equipment. The protective earth (ground) terminal must be bonded to protective earth in the end use equipment.

#### **CONNECTIONS / DIMENSIONS:**



AIR FLOW DIRECTION:

1-Optimum, 2- Good, 3- Fair

**CONNECTORS:** 

P1: DC Input - .156 friction lock header mates with TE Connectivity 640250-4 or equivalent crimp terminal housing with TE Connectivity 3-640706-1 or equivalent crimp terminal

P2: DC Output (Single and Multiple) - .156 friction lock header mates with TE Connectivity 770849-8 or equivalent crimp terminal housing with TE Connectivity 3-640707-1 or equivalent crimp terminal.

G: Ground- .187 quick disconnect terminal.

P3: P.G./Sense (Single) - .100 breakaway header mates with Molex 22-55-2061 or equivalent crimp terminal housing with Molex type 71851 or equivalent crimp terminal.

P3: Power Good (Multi) - .100 breakaway header mates with Molex 50-57-9002 or equiv. crimp terminal housing with Molex type 71851 or equiv. crimp terminal.