300 Stewart Road Wilkes-Barre, PA 18706 USA Phone: (570) 824-4666 Web: www.ipdpower.com Email: sales@ipdpower.com Fax: (570) 824-4843

REL-185 SERIES OPERATING INSTRUCTIONS

INPUT RATING: 100-240VAC, 4 A, 50-60 Hz.

OUTPUT RATING: 185 Watts Maximum Total Continuous Output Power with 300 LFM Forced Air.

135 Watts Maximum Total Continuous Output Power Convection Cooled.

110 Watts Maximum Total Continuous Output Power Convection Cooled with Chassis and Cover.

MODEL LISTING:

<u>Model</u>	Output 1 ₍₁₎	Output 2(1)	Output 3 ₍₂₎	Output 4 ₍₂₎
REL-185-4001	+3.3VDC/20A(3)	+5VDC/10A	+12VDC/2A	-12VDC/2A
REL-185-4002	+5VDC/20A(3)	+3.3VDC/10A	+12VDC/2A	-12VDC/2A
REL-185-4003	+5VDC/20A(3)	+3.3VDC/10A	+15VDC/2A	-15VDC/2A
REL-185-4004	+5VDC/20A(3)	-5VDC/10A	+12VDC/2A	-12VDC/2A
REL-185-4005	+5VDC/20A(3)	-5VDC/10A	+15VDC/2A	-15VDC/2A
REL-185-4006	+5VDC/20A(3)	+24VDC/3A	+12VDC/2A	-12VDC/2A
REL-185-4007	+5VDC/20A(3)	+24VDC/3A	+15VDC/2A	-15VDC/2A
REL-185-4008	+3.3VDC/20A(3)	+6.7VDC/5A	+12VDC/2A	-6.7VDC/2A
REL-185-4009	5VDC/15A	12VDC/4A	28VDC/0.8A ₍₆₎	12VDC/2A
REL-185-4010	5VDC/20A	5VDC/10A	$24VDC/1A_{(6)}$	24VDC/1A(6)
REL-185-3001	+5VDC/20A(3)	+12VDC/5A		-12VDC/3A
REL-185-3002	+5VDC/20A(3)	+15VDC/4A		-15VDC/3A
REL-185-2001	+3.3VDC/20A(3)	+5VDC/10A		
REL-185-2002	+5VDC/20A(3)	+12VDC/8A		
REL-185-2003	+5VDC/20A(3)	+24VDC/4A		
REL-185-2004	+12VDC/10A	-12VDC/4A		
REL-185-2005	+15VDC/8A	-15VDC/5A		
REL-185-2006	+15VDC/6A	+24VDC/4A		
REL-185-2007	+35VDC/3.5A	+12VDC/5.2A		
REL-185-1001	2.5VDC/37A(4)			
REL-185-1002	3.3VDC/37A(4)			
REL-185-1003	5VDC/37A(4)			
REL-185-1004	12VDC/15.4A			
REL-185-1005	15VDC/12.3A			
REL-185-1006	24VDC/7.7A			
REL-185-1007	28VDC/6.6A			
REL-185-1008	48VDC/3.8A			
REL-185-1009	6.3VDC/29A			
REL-185-1010	31VDC/6A			

NOTES:

- 1. Total current from Outputs 1 and 2 must not exceed 20 Amps with convection cooling.
- 2. Total current from Outputs 3 and 4 must not exceed 3 Amps with convection cooling.
- 3. Rated 15 Amps maximum with convection cooling.
- 4. Rated 27 Amps maximum with convection cooling.
- 5. A suffix may be added to the model number to indicate the following optional configurations: (CH-chassis, CO-cover, IO-isolated outputs, RE-remote inhibit, TS-terminal strip)

6. Output is ES3 type.

CE

DECLARATION OF CONFORMITY



Manufacturer: Integrated Power Designs, Inc.

Manufacturer's Address: 300 Stewart Road, Wilkes-Barre, PA 18706 USA

Product Type: Switchmode AC-DC Power Supply

Declares all models listed above including all options are in compliance with the following European Community Directives:

Low Voltage Directive 2014/35/EU of 26 February 2014 RoHS Directive 2015/863/EU of 31 March 2015

As well as with the requirements set out in the Laws of the United Kingdom relating to and applying the following standards:

Electrical Equipment (Safety) Regulations 2016 SI No. 1101

BS EN 62368-1:2014

Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 SI No. 3032 + 2019 SI No.492

In addition, all models are Certified to be in compliance with applicable requirements of UL 62368-1:2014, ANSI/AAMI ES60601:2005/(R) 2012, IEC 62368-1:2014 and IEC 60601-1:2005/A1:2012 including all EU national deviations, CAN/CSA-C22.2 No. 62368-1-14, CAN/CSA-C22.2 No. 60601-1:2014, EN 62368-1:2014 and EN 60601-1:2006/A1:2013.

BY: Steven Thompson- President

PLACE: Integrated Power Designs

300 Stewart Road, Wilkes-Barre, PA 18706 USA

DATE: February 16, 2021

EUROPEAN CONTACT:

Compumess Elektronik GmbH Lise-Meitner-Strasse 1 85716 Unterschleiβheim Telephone (089) 32 15 01-0 **UK CONTACT:**

ALL PSU LTD Laser Quay, Culpeper Close Rochester Kent ME2 4HU Tel: 01634 725527



300 Stewart Road Wilkes-Barre, PA 18706 USA Phone: (570) 824-4666

Web: www.ipdpower.com Email: sales@ipdpower.com Fax: (570) 824-4843

CLASSIFICATION:

- 1. Protection against electric shock Class I.
- 2. Protection against harmful ingress of water IPXO (Non-protected), ordinary.
- 3. Methods of sterilization None.
- 4. Suitability for use in an oxygen rish environment End user responsibility, not evaluated.
- 5. Mode of operation Continuous.

WARNING! RISK OF FIRE! An open internal fuse indicates a catastrophic failure of circuit component(s). Repair must be by authorized IPD personnel only. Refer to fuse rating on power supply circuit board for rating.

WARNING! SHOCK HAZARD! Dangerous voltages are present on some components, printed circuit board traces and heatsinks.

This product includes a single fuse in the phase lead only. In consideration of IEC 60601-1:2005 Clause 8.11.5, a second fuse may be required in the neutral lead of the end use equipment

SEPERATION:

INPUT FUSE:

Primary to secondary creepage distance is 8mm minimum, clearance 5mm minimum (2MOPP). Primary to ground creepage distance is 4mm minimum, clearance 2.5mm minimum (1MOPP). Secondary to ground creepage is 2.3mm minimum, clearance is 1.4mm minimum (Operational Insulation). Required seperation must be maintained in the end use equipment to preserve the established means of protection.

OUTPUTS:

The outputs are not acceptable for patient connection without additional isolation. All outputs are ES1 under normal and single fault conditions unless otherwise indicated.

TEMPERATURES:

The maximum operating temperatures of safety components as defined in the applicable safety standards must not be exceeded after installation in the end use equipment. Output power, ambient air temperature and convection or forced air cooling availability should be considered in the end use equipment.

HIPOT:

In consideration of IEC 60601-1:2005 Clause 8.8.3, care must be taken to insure the voltage applied to a reinforced insulation does not overstress different types and levels of insulation. Breakdown of basic insulation and catastrophic failure of the power supply may result if a test voltage of greater than 1800 VAC is applied between primary and secondary circuits. Each isolating component is factory tested at 4000 VAC minimum prior to installation. Primary and secondary to ground capacitors may need to be disconnected prior to performing a dielectric strength test on the power supply or the end product.

INSTALLATION:


The power supplies included in model listing on reverse side are considered components intended for professional installation into end use equipment. The protective earth (ground) terminal must be bonded to protective earth in the end use equipment.

EMISSIONS:

This product was tested for compliance with EN 55022 and EN 55011 Class B conducted and radiated emissions using the techniques listed below and non-inductive load resistors to simulate operation in a typical installation. All or a combination of the following requirements may be necessary to insure compliance in the end use equipment.

- 1. Installation of the power supply, output cables and loads in a shielded enclosure.
- 2. Use of optional chassis and cover.
- 3. Use of shielded I/O cables.
- 4. Use of ferrite beads on I/O cables.
- 5. Grounded output returns.

CONNECTIONS / DIMENSIONS:



AIR FLOW **DIRECTION:**

CONNECTORS:

- AC Input 0.156 inch friction lock header mates with Molex 09-50-3031 or equivalent crimp terminal housing with Molex 2478 or equivalent crimp terminal.
- DC Output (Single Output) 6-32 screw-down terminal mates with #6 ring-tongue terminal (10 in.-lb. max.).
- DC Output (Multiple Output) 0.156 inch friction lock header mates with Molex 09-50-3161 or equivalent crimp terminal housing with Molex 2478 or equivalent crimp terminal.
- Protective Earth (Ground) 0.187 quick disconnect terminal.
- Option / Sense (Single Output) 0.100 friction lock header mates with Molex 50-57-9008 or equivalent crimp terminal housing with Molex type 71851 or equivalent crimp terminal.
- Option / Sense (Multiple Output) 0.100 breakaway header mates with Molex 22-55-2081 or equivalent crimp terminal housing with Molex type 71851 or equivalent crimp terminal.